

Model CB 260 | CO₂ incubators with hot air sterilization and heat sterilizable CO₂ sensor

The BINDER CB series CO₂ incubator is the premium class among the CO₂ incubators. It is suitable for all sensitive incubation applications and ensures optimal cell growth. A BINDER CB series CO₂ incubator can perform complex cultivation experiments as well as specific environmental conditions.

BENEFITS

- Maximum usable space with minimal footprint and excellent ease of use
- Minimum risk of contamination thanks to uncompromising hot-air sterilization
- Fanless, easy-to-clean hygienic interior without fixtures
- Stable pH values thanks to drift-free CO₂ IR sensor technology
- Low evaporation of the medium due to high levels of humidity and quick recovery time



Model 260



Model 260

MAIN FEATURES

- Temperature range: from 4 °C above ambient temperature up to 60 °C
- Humidity range: 90 to 95% RH
- CO₂ control range: 0 to 20 vol.% CO₂
- Tried-and-tested anti-contamination concept with 180 °C hot air sterilization
- Optimized double-pan humidification system with condensation protection
- CO₂ gas-mixing jet with Venturi effect
- Hot-air sterilizable CO₂ sensor with single-beam infrared technology
- Intuitive touchscreen controller
- Internal data recording, can be read out in open format via USB interface
- Fail-safe CO₂ system for protecting the pH of cell cultures
- DuoDoor™ system: Unit door and glass door can be opened with one handle
- Tightly fitted inner door made of safety glass
- Lockable door handle
- Seamless inner chamber made of stainless steel with flanges as shelf support system
- Perforated shelves, stainless steel
- Troubleshooting system with visual and acoustic alarms
- Interfaces: Ethernet, USB, zero-voltage alarm contact
- Units can be stacked with CB-S models of the same size using a stacking adapter

OPTIONAL EQUIPMENT

- Versions with O₂ control range: 0.2 to 20 vol.% O₂ or 10 to 95 vol.% O₂
- Versions with humidity control range: 50 to 95% RH due to active humidification
- Convenient water supply with reusable tubular bags, pre-filled single-liter water bottle, or water container set with magnetic holder, for versions with active humidification
- Multi-sectional glass doors with special shelves for dividing up the space to accommodate different experiments
- Stacking adapter for secure, space-saving installation
- CO₂ gas tank changer for connecting two gas tanks to one or two incubators
- CO₂ shaker with corrosion-resistant components in a stainless-steel housing
- APT-COM Multi Management Software – for managing, recording, and documenting unit parameters
- Services – wide range of services to ensure that the unit functions correctly

ORDERING INFORMATION

Interior volume [cu.ft.]	Voltage	Option model	Version	Art.-No.
9.4	200...230 V 1~ 50/60 Hz	Standard	CB260-230V	9040-0147
		with O ₂ control	CB260-230V-O	9040-0148
		with active humidification	CB260-230V-F	9040-0149
		with O ₂ control 0.2–20% and active humidification	CB260-230V-OF	9040-0150
	100...120 V 1~ 50/60 Hz	Standard	CB260UL-120V	9040-0152
		with O ₂ control	CB260UL-120V-O	9040-0153
		with active humidification	CB260UL-120V-F	9040-0154
		with O ₂ control 0.2–20% and active humidification	CB260UL-120V-OF	9040-0155

TECHNICAL DATA (Version 1-4 from 8)

Description	CB260-230V ¹	CB260-230V-O ¹	CB260-230V-F ¹	CB260-230V-OF ¹
Article Number	9040-0147	9040-0148	9040-0149	9040-0150
Performance Data Temperature				
Temperature range +5 °C above ambient temperature to	60	60	60	60
Temperature variation at 37 °C [± K]	0.4	0.4	0.4	0.4
Temperature fluctuation at 37 °C [± K]	0.1	0.1	0.1	0.1
Recovery time after 30 seconds door open at 37 °C [min]	6	6	6	6
Performance Data Climate				
Humidity range [% RH]	90...95	90...95	90...95	90...95
CO₂-Data				
CO ₂ range [Vol.-% CO ₂]	0...20	0...20	0...20	0...20
CO ₂ measurement	IR	IR	IR	IR
CO ₂ Recovery time after 30 seconds door open at 5 Vol.-% CO ₂ [min]	5	5	5	5
O₂-Data				
Standard O ₂ control ranges: O ₂ range [Vol.-% O ₂]		0.2...20		0.2...20
O ₂ control ranges with option: O ₂ range [Vol.-% O ₂]		10...95		10...95
O ₂ Recovery time after 30 seconds door open at 5 Vol.-% O ₂ [min]		15		15
Electrical data				
Rated Voltage [V]	200...230	200...230	200...230	200...230
Power frequency [Hz]	50/60	50/60	50/60	50/60
Nominal power [kW]	1.5	1.5	1.5	1.5
Unit fuse [A]	10	10	10	10
Phase (Nominal voltage)	1~	1~	1~	1~
Measures				
Interior volume [cu.ft.]	9.4	9.4	9.4	9.4
Net weight of the unit (empty) [lbs]	276	276	276	276
Permitted load [lbs]	88	88	88	88
Load per rack [lbs]	22	22	22	22
Wall clearance back [in]	3.94	3.94	3.94	3.94
Wall clearance sidewise [in]	1.97	1.97	1.97	1.97

1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

Description	CB260-230V ¹	CB260-230V-0 ¹	CB260-230V-F ¹	CB260-230V-OF ¹
Article Number	9040-0147	9040-0148	9040-0149	9040-0150
Internal Dimensions				
Width [in]	24.43	24.43	24.43	24.43
Height [in]	29.55	29.55	29.55	29.55
Depth [in]	22.65	22.65	22.65	22.65
Doors				
Inner doors	1	1	6	6
Unit doors	1	1	1	1
Housing dimensions not incl. fittings and connections				
Width net [in]	29.16	29.16	29.16	29.16
Height net [in]	40.2	40.2	40.2	40.2
Depth net [in]	30.93	30.93	30.93	30.93
Environment-specific data				
Energy consumption at 37 °C [Wh/h]	65	65	65	65
Sound-pressure level [dB(A)]	41	41	41	41
Fixtures				
Number of shelves (std./max.)	3/8	3/8	3/8	3/8

- 1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

TECHNICAL DATA (Version 5-8 from 8)


Description	CB260UL-120V ¹	CB260UL-120V-0 ¹	CB260UL-120V-F ¹	CB260UL-120V-OF ¹
Article Number	9040-0152	9040-0153	9040-0154	9040-0155
Performance Data Temperature				
Temperature range +5 °C above ambient temperature to	60	60	60	60
Temperature variation at 37 °C [± K]	0.4	0.4	0.4	0.4
Temperature fluctuation at 37 °C [± K]	0.1	0.1	0.1	0.1
Recovery time after 30 seconds door open at 37 °C [min]	6	6	6	6
Performance Data Climate				
Humidity range [% RH]	90...95	90...95	90...95	90...95
CO ₂ -Data				
CO ₂ range [Vol.-% CO ₂]	0...20	0...20	0...20	0...20
CO ₂ measurement	IR	IR	IR	IR
CO ₂ Recovery time after 30 seconds door open at 5 Vol.-% CO ₂ [min]	5	5	5	5
O ₂ -Data				
Standard O ₂ control ranges: O ₂ range [Vol.-% O ₂]		0.2...20		0.2...20
O ₂ control ranges with option: O ₂ range [Vol.-% O ₂]		10...95		10...95
O ₂ Recovery time after 30 seconds door open at 5 Vol.-% O ₂ [min]		15		15

- 1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

Description	CB260UL-120V ¹	CB260UL-120V-0 ¹	CB260UL-120V-F ¹	CB260UL-120V-OF ¹
Article Number	9040-0152	9040-0153	9040-0154	9040-0155
Electrical data				
Rated Voltage [V]	100...120	100...120	100...120	100...120
Power frequency [Hz]	50/60	50/60	50/60	50/60
Nominal power [kW]	1.5	1.5	1.5	1.5
Unit fuse [A]	16	16	16	16
Phase (Nominal voltage)	1~	1~	1~	1~
Measures				
Interior volume [cu.ft.]	9.4	9.4	9.4	9.4
Net weight of the unit (empty) [lbs]	276	276	276	276
Permitted load [lbs]	88	88	88	88
Load per rack [lbs]	22	22	22	22
Wall clearance back [in]	3.94	3.94	3.94	3.94
Wall clearance sidewise [in]	1.97	1.97	1.97	1.97
Internal Dimensions				
Width [in]	24.43	24.43	24.43	24.43
Height [in]	29.55	29.55	29.55	29.55
Depth [in]	22.65	22.65	22.65	22.65
Doors				
Inner doors	1	1	6	6
Unit doors	1	1	1	1
Housing dimensions not incl. fittings and connections				
Width net [in]	29.16	29.16	29.16	29.16
Height net [in]	40.2	40.2	40.2	40.2
Depth net [in]	30.93	30.93	30.93	30.93
Environment-specific data				
Energy consumption at 37 °C [Wh/h]	65	65	65	65
Sound-pressure level [dB(A)]	41	41	41	41
Fixtures				
Number of shelves (std./max.)	3/8	3/8	3/8	3/8

1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

DIMENSIONS INCL. FITTINGS AND CONNECTIONS [MM]



OPTIONS

Designation	Description	*	Art.-No.
Calibration certificate, temperature	temperature measurement incl. certificate and 27 measuring points at specified temperature	-	8012-1592
	8-pin, for low voltage with LEMO socket (coverable) and LEMO plug (max. 24 V – 2 A)	-	8012-1848
	alternative to the standard range, for controlling O ₂ concentrations between 10 and 95 vol. %, only in conjunction with option models with O ₂ control	-	9419-0134
	back		
	30 mm	01	8012-1390
	for temperature and CO ₂ values (outputs not adjustable)	02	8012-1839
	for temperature and CO ₂ values (outputs not adjustable), for versions with O ₂ control	02	8012-1840
	for temperature and CO ₂ , temperature measurement in center of chamber, CO ₂ measurement performed using test gas at 37 °C and 5 % CO ₂	-	8012-1235
	for temperature; for extending the measurement in center of chamber to include another test temperature	-	8012-1113
	for versions with O ₂ control; a measuring point in center of chamber at a specified O ₂ percentage value	-	8012-0229
	Internal door divided into 8 parts		
	door hinged on the left (incl. option door hinged on the left)	-	8012-1926
Options/Accessories	door hinged on the right	-	8012-1845
	left		
	30 mm	01	8012-1391
	Measurement in center of chamber at specified temperature	-	8012-1132
	modular, expandable roller bottle system for cell cultivation, 6 roller bottles	-	8012-0572
	right		
	30 mm	01	8012-1392
	temperature measurement incl. certificate, 15- 18 measuring points at specified temperature	-	8012-1571
	temperature measurement incl. certificate, 9 measuring points at specified temperature	-	8012-1550
	Unit door and standard glass inner door with door hinged on the left	-	8012-1915
	viewing window (13 x 25 cm) in the inner door for quick access		
	door hinged on the left (incl. option door hinged on the left)	-	8012-1925
	door hinged on the right	-	8012-1923

* Notes › See last page

ACCESSORIES

Designation	Description	*	Art.-No.
	Multi Management Software APT-COM™		
Multi Management Software	version 4, BASIC edition	19	9053-0039
	version 4, GLP edition	19	9053-0042
	version 4, PROFESSIONAL edition	19	9053-0040

* Notes › See last page

Designation	Description	*	Art.-No.
	Stainless steel	-	6004-0197
	concentrated, for gentle remove of residual contaminants; 1 kg	-	1002-0016
	external, BINDER Gas Supply Service, for connecting 2 gas tanks (CO ₂ , N ₂ or O ₂), with audible and visual alarms, as well as zero-voltage alarm output	-	8012-0408
	for CO ₂ , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	-	8012-0014
	for divided inner door		
	Stainless steel, divided	-	8012-1996
	for N ₂ , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	-	8012-0016
	for O ₂ , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	-	8012-0015
	for the thermally isolated stacking of two BINDER CO ₂ incubators	-	9051-0039
	IQ/OQ documents – supporting documents for validation performed by customers, consisting of: IQ/OQ checklists incl. calibration guide and comprehensive unit documentation; parameters: temperature, CO ₂ , O ₂ , pressure, depending on unit		
	Digital in PDF format	-	7057-0001
	Hard copy inside folder	-	7007-0001
	IQ/OQ/PQ documents – supporting documents for validation performed by customers, according to customer requirements, PQ section added to qualification folder IQ/OQ; parameters: temperature, CO ₂ , O ₂ – or pressure, depending on unit		
Options/Accessories	Digital in PDF format	-	7057-0005
	Hard copy inside folder	-	7007-0005
	Set consisting of 10 bottles, each with 1,000 ml of sterile humidification water, including peristaltic tubing, connectors, and fine gas filters		
	For versions with active humidification	-	8009-1104
	Set consisting of 10 liter water container with magnetic holder, hose adapter with practical quick coupling and water stop function, flexible positioning, enables water to be supplied to multiple CB units simultaneously via a water container, not pre-filled.		
	For versions with active humidification	-	8009-1128
	Set consisting of 12 empty, refillable tubular bags including peristaltic tubing, connectors, and fine gas filters		
	For versions with active humidification	-	8009-1058
	Set consisting of 3 empty, refillable tubular bags including peristaltic tubing, connectors, and fine gas filters		
	For versions with active humidification	-	8009-1083
	set of 4 anti-slip feet, for safe stacking of units	-	8012-0702
	Shaker and controller connected via ribbon cable, universally applicable, 100 – 240 V, 50/60 Hz	-	8012-1771
	Shaker and controller equipped with LEMO connectors, exclusively for CB with electric access port, 100 – 240 V, 50/60 Hz	-	8012-1772
	stable, vibration-free stacking frame on casters with locking brake for the safe stacking of two CO ₂ incubators	-	9051-0041
	the base equipped with casters is used for the secure positioning and leveling of a BINDER CO ₂ incubator	-	9051-0044

* Notes › See last page

SERVICES

Designation	Description	*	Art.-No.
Installation services			
Services	Unit function instructions for operation and programming of the controller	18	DL10-0500
	Unpacking and setting up of unit, connecting to existing connections, and checking function	13, 18	DL10-0100
Preventive maintenance			
Services	Executive of equipment inspection according to maintenance plan	14, 18	DL20-0400

* Notes › See last page

Designation	Description	*	Art.-No.
Calibration services			
	CO ₂ calibration at a specified value, measurements are taken with analyzed test gas at 5%, including certificate	14, 16, 17, 18	DL30-0401
	Expansion – including certificate, each additional measuring point in center of chamber at specified test temperature	14, 16, 17, 18	DL30-0102
	O ₂ calibration with 1 measuring point in center of chamber at a specified O ₂ percentage value, including certificate	14, 16, 17, 18	DL30-0402
Services	Temperature calibration with 1 measuring point in center of chamber at a specified test temperature, including certificate	14, 16, 17, 18	DL30-0101
	Temperature measurement with 18 measuring points at a specified test temperature, including certificate	14, 16, 17, 18	DL30-0118
	Temperature measurement with 27 measuring points at a specified test temperature, including certificate	14, 16, 17, 18	DL30-0127
	Temperature measurement with 9 measuring points at a specified test temperature, including certificate	14, 16, 17, 18	DL30-0109
Validation services			
Services	Execution of IQ/OQ in accordance with qualification folder	15, 18	DL41-0200
	Execution of IQ/OQ/PQ in accordance with qualification folder	15, 18	DL44-0500
Warranty service			
Services	The warranty is extended from 2 to 3 years from the delivery date, wear parts are excluded	–	DL02-1041
	The warranty is extended from 2 to 5 years from the delivery date, wear parts are excluded	–	DL02-1042

* Notes › See last page

NOTES

- 01 Condensation may occur in the area around the access port. Access ports may be placed in custom locations for an additional charge.
- 02 UL mark is not granted when this option is used.
- 03 Heat resistant only to max. 200 °C.
- 04 Only available on units rated for 230 V.
- 07 The additional heat input may affect the temperature behavior.
- 09 Not in conjunction with the optional access port, door with window and interior lighting.
- 10 Not available on 23-liter units.
- 11 Not available on 23- or 53-liter units.
- 12 Only available on units rated for 230 V or 400 V.
- 13 Installation and connections take place at unit location; transport within the company only upon consultation.
- 14 We recommend a BINDER service contract to cover unit inspections, calibrations and validations.
- 15 OQ according to Yellow Paper = completed factory validation documentation of all OQ checklists.
- 16 Sensor calibration is performed in an accredited calibration laboratory.
- 17 Calibration is performed according to the BINDER factory standard.
- 18 Quoted prices do not include travel costs. Please refer to the chapter on BINDER Service for travel costs for your region. Quoted prices for services performed in Switzerland do not include a country-specific added fee (available on request).
- 19 For additional accessories, refer to the Process documentation chapter.

BINDER GmbH

Tuttlingen, Germany
TEL +49 7462 2005 0
FAX +49 7462 2005 100
info@binder-world.com
www.binder-world.com

BINDER Asia Pacific (Hong Kong) Ltd.

Kowloon, Hong Kong, P.R. China
TEL +852 39070500
FAX +852 39070507
asia@binder-world.com
www.binder-world.com

Representative Office for CIS

Moscow, Russia
TEL +7 495 988 15 16
FAX +7 495 988 15 17
russia@binder-world.com
www.binder-world.com

**BINDER Environmental Testing
Equipment (Shanghai) Co., Ltd.**

Shanghai, P.R. China
TEL +86 21 685 808 25
FAX +86 21 685 808 29
china@binder-world.com
www.binder-world.com

BINDER Inc.

Bohemia, NY, USA
TEL +1 631 224 4340
FAX +1 631 224 4354
usa@binder-world.com
www.binder-world.us